



Ingredients: Pure water + Ethanol Solution (<1%)

Life: one year.

The life is approx. 1 year if the bottle is covered when it doesn't work.

Function: to verify the condition of the unit and sensor.

Operation:

- Set the sensitivity level to "Hi".
- Open the leak check bottle cover and slowly move it closer to snake tube nozzle.
- If the indication moves up to high from low then we should move the check bottle away and the LED should go off again. This shows that the unit is under working condition.
- If the unit does not perform as we expect, bring the unit for maintenance at your local sales office.

Note/remark:

If the Leak Check Bottle is dry, it cannot work.

The bottle must be closed if the users don't use it.

Please try to test the detectors with new leak check bottle.